

BFO

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 09/14/2021

OPERATOR NAME:  
ROBERT S CORNISH  
Permit #: 48068  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:27
INTERNAL STANDARD	VERIFIED	21:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:27
BLANK TEST	0.00	21:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:29
BLANK TEST	0.00	21:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:30
BLANK TEST	0.00	21:31

