

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 10/04/2021

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:20
INTERNAL STANDARD	VERIFIED	23:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:21
BLANK TEST	0.00	23:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:22
BLANK TEST	0.00	23:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:23
BLANK TEST	0.00	23:24

