

BFO

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/03/2021

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:09
INTERNAL STANDARD	VERIFIED	03:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:10
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:11
BLANK TEST	0.00	03:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:12
BLANK TEST	0.00	03:13

