

**SUPERVISOR TEST**

-----  
 ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 141906  
 -----

Date: 07/21/2021

OPERATOR NAME:  
 STEVEN P EILERS  
 Permit #: 49321  
 Operator Agency/Dept: ONEIDA CO. S.O.  
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:48
INTERNAL STANDARD	VERIFIED	21:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:48
BLANK TEST	0.00	21:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:50
BLANK TEST	0.00	21:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:51
BLANK TEST	0.00	21:52

