

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 10/13/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:53
INTERNAL STANDARD	VERIFIED	23:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:54
BLANK TEST	0.00	23:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:55
BLANK TEST	0.00	23:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:56
BLANK TEST	0.00	23:57

