

BFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/08/2021

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:51
INTERNAL STANDARD	VERIFIED	21:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:52
BLANK TEST	0.00	21:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:53
BLANK TEST	0.00	21:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:54
BLANK TEST	0.00	21:55

