

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/22/2021

OPERATOR NAME:
JAMES P QUINN
Permit #: 49340
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:05
INTERNAL STANDARD	VERIFIED	20:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:05
BLANK TEST	0.00	20:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:06
BLANK TEST	0.00	20:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:07
BLANK TEST	0.00	20:08

BFO

Weekly Data Master
Supervisory test

12/22/21

