

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/29/2021

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:20
INTERNAL STANDARD	VERIFIED	18:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:20
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:22
BLANK TEST	0.00	18:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:23
BLANK TEST	0.00	18:24

