

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 05/02/2022

OPERATOR NAME:  
KRISTY N ALLEN  
Permit #: 49726  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:55
INTERNAL STANDARD	VERIFIED	01:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:55
BLANK TEST	0.00	01:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:56
BLANK TEST	0.00	01:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:58
BLANK TEST	0.00	01:58

