

SOLUTION CHANGE

 ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 141906

Date: 12/01/2022

OPERATOR NAME:
 KRISTY N ALLEN
 Permit #: 49726
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:34
INTERNAL STANDARD	VERIFIED	05:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:34
BLANK TEST	0.00	05:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:35
BLANK TEST	0.00	05:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:37
BLANK TEST	0.00	05:38

