

BFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 05/11/2022

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:53
INTERNAL STANDARD	VERIFIED	17:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:53
BLANK TEST	0.00	17:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:55
BLANK TEST	0.00	17:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:56
BLANK TEST	0.00	17:57

