

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 03/29/2022

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:35
INTERNAL STANDARD	VERIFIED	23:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:35
BLANK TEST	0.00	23:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:37
BLANK TEST	0.00	23:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:38
BLANK TEST	0.00	23:39

