

BFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 03/30/2022

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:54
INTERNAL STANDARD	VERIFIED	18:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:55
BLANK TEST	0.00	18:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:56
BLANK TEST	0.00	18:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:57
BLANK TEST	0.00	18:58

