

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 04/14/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:28
INTERNAL STANDARD	VERIFIED	21:28
REFERENCE STANDARD TEMP	33.99c	.
REFERENCE STANDARD	0.10	21:29
BLANK TEST	0.00	21:30
REFERENCE STANDARD TEMP	33.99c	.
REFERENCE STANDARD	0.10	21:30
BLANK TEST	0.00	21:31
REFERENCE STANDARD TEMP	33.99c	.
REFERENCE STANDARD	0.10	21:31
BLANK TEST	0.00	21:32

