

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 04/20/2022

OPERATOR NAME:  
KRISTY N ALLEN  
Permit #: 49726  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:15
INTERNAL STANDARD	VERIFIED	18:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:16
BLANK TEST	0.00	18:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:17
BLANK TEST	0.00	18:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:18
BLANK TEST	0.00	18:19

