

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 05/19/2022

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:35
INTERNAL STANDARD	VERIFIED	01:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:36
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:37
BLANK TEST	0.00	01:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:38
BLANK TEST	0.00	01:39

