

BFO

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 05/25/2022

OPERATOR NAME:  
LEONARD F DILLENBECK  
Permit #: 47621  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:52
INTERNAL STANDARD	VERIFIED	17:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:52
BLANK TEST	0.00	17:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:53
BLANK TEST	0.00	17:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:55
BLANK TEST	0.00	17:56

