

SUPERVISOR TEST

BF0

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 07/20/2022

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:31
BLANK TEST	0.00	03:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:32
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:34
BLANK TEST	0.00	03:34

