

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 08/17/2022

OPERATOR NAME:  
KRISTY N ALLEN  
Permit #: 49726  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:26
INTERNAL STANDARD	VERIFIED	00:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:27
BLANK TEST	0.00	00:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:28
BLANK TEST	0.00	00:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:29
BLANK TEST	0.00	00:30

