

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 09/28/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:39
INTERNAL STANDARD	VERIFIED	00:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:40
BLANK TEST	0.00	00:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:42
BLANK TEST	0.00	00:43

