

SUPERVISOR TEST

Kfo

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/09/2022

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:44
INTERNAL STANDARD	VERIFIED	01:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:45
BLANK TEST	0.00	01:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:46
BLANK TEST	0.00	01:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:47
BLANK TEST	0.00	01:48

