

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFD

Date: 11/23/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:19
INTERNAL STANDARD	VERIFIED	02:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:20
BLANK TEST	0.00	02:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:21
BLANK TEST	0.00	02:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:22
BLANK TEST	0.00	02:23

