

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/07/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:06
INTERNAL STANDARD	VERIFIED	01:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:08
BLANK TEST	0.00	01:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:09
BLANK TEST	0.00	01:10

BFO

