

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 05/03/2023

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23030

*BFO  
monthly*

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:14
INTERNAL STANDARD	VERIFIED	00:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:15
BLANK TEST	0.00	00:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:16
BLANK TEST	0.00	00:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:17
BLANK TEST	0.00	00:18

