

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 08/01/2023

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:50
INTERNAL STANDARD	VERIFIED	04:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	04:51
BLANK TEST	0.00	04:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	04:52
BLANK TEST	0.00	04:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	04:53
BLANK TEST	0.00	04:54

