

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/01/2023

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:25
INTERNAL STANDARD	VERIFIED	22:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:25
BLANK TEST	0.00	22:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:27
BLANK TEST	0.00	22:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:28
BLANK TEST	0.00	22:29

