

SOLUTION CHANGE

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 03/01/2023

OPERATOR NAME:  
KRISTY N ALLEN  
Permit #: 49726  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:38
INTERNAL STANDARD	VERIFIED	00:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:38
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:40
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:41
BLANK TEST	0.00	00:42

