

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/26/2023

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 03:52 |
| INTERNAL STANDARD | VERIFIED | 03:52 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 03:52 |
| BLANK TEST | 0.00 | 03:53 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 03:53 |
| BLANK TEST | 0.00 | 03:54 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 03:55 |
| BLANK TEST | 0.00 | 03:56 |

