

SUPERVISOR TEST

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 05/24/2023

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:46
INTERNAL STANDARD	VERIFIED	05:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:47
BLANK TEST	0.00	05:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:48
BLANK TEST	0.00	05:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:49
BLANK TEST	0.00	05:50

