

SUPERVISOR TEST

BFB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 05/31/2023

OPERATOR NAME:  
KRISTY N ALLEN  
Permit #: 49726  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:41
INTERNAL STANDARD	VERIFIED	22:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:41
BLANK TEST	0.00	22:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:43
BLANK TEST	0.00	22:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:44
BLANK TEST	0.00	22:45

