

SUPERVISOR TEST

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 07/26/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:16
INTERNAL STANDARD	VERIFIED	23:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:17
BLANK TEST	0.00	23:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:18
BLANK TEST	0.00	23:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:19
BLANK TEST	0.00	23:20

