

SUPERVISOR TEST

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 08/16/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:26
INTERNAL STANDARD	VERIFIED	02:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:26
BLANK TEST	0.00	02:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:28
BLANK TEST	0.00	02:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:29
BLANK TEST	0.00	02:30

