

BFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/11/2023

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:39
INTERNAL STANDARD	VERIFIED	01:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:40
BLANK TEST	0.00	01:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:41
BLANK TEST	0.00	01:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:42
BLANK TEST	0.00	01:43

