

BFO

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 02/08/2023

OPERATOR NAME:  
LEONARD F DILLENBECK  
Permit #: 47621  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22420\

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:18
INTERNAL STANDARD	VERIFIED	06:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:18
BLANK TEST	0.00	06:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:20
BLANK TEST	0.00	06:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:21
BLANK TEST	0.00	06:22

