

SUPERVISOR TEST

BFO

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 141906

Date: 02/15/2023

OPERATOR NAME:
 KRISTY N ALLEN
 Permit #: 49726
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:07
INTERNAL STANDARD	VERIFIED	03:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:07
BLANK TEST	0.00	03:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:08
BLANK TEST	0.00	03:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:09
BLANK TEST	0.00	03:10

