

SUPERVISOR TEST

BFO

ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 141906

Date: 03/08/2023

OPERATOR NAME:  
 JOHN P CUSACK  
 Permit #: 35563  
 Operator Agency/Dept: ONEIDA CO. S.O.  
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:33
INTERNAL STANDARD	VERIFIED	22:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:34
BLANK TEST	0.00	22:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:35
BLANK TEST	0.00	22:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:36
BLANK TEST	0.00	22:37

