

SOLUTION CHANGE

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 01/31/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:30
INTERNAL STANDARD	VERIFIED	02:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:31
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:32
BLANK TEST	0.00	02:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:33
BLANK TEST	0.00	02:34

