

SOLUTION CHANGE

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 03/01/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:34
INTERNAL STANDARD	VERIFIED	23:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:34
BLANK TEST	0.00	23:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:36
BLANK TEST	0.00	23:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:37
BLANK TEST	0.00	23:38

