

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

BFO

Date: 12/02/2024

OPERATOR NAME:  
NATHAN R VALENTINE  
Permit #: 53025  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:13
INTERNAL STANDARD	VERIFIED	02:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:14
BLANK TEST	0.00	02:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:15
BLANK TEST	0.00	02:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:16
BLANK TEST	0.00	02:17

