

SOLUTION CHANGE

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 01/01/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:50
INTERNAL STANDARD	VERIFIED	22:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:50
BLANK TEST	0.00	22:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:51
BLANK TEST	0.00	22:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:52
BLANK TEST	0.00	22:53

