

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/02/2024

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:01
INTERNAL STANDARD	VERIFIED	18:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:01
BLANK TEST	0.00	18:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:03
BLANK TEST	0.00	18:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:04
BLANK TEST	0.00	18:05

BFO

