

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/17/2024

OPERATOR NAME:
ROBERT J PRODY
Permit #: 50912
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:37
INTERNAL STANDARD	VERIFIED	22:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:39
BLANK TEST	0.00	22:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:40
BLANK TEST	0.00	22:41

BFO

