

SUPERVISOR TEST

BFD

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 04/23/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:54
INTERNAL STANDARD	VERIFIED	21:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:55
BLANK TEST	0.00	21:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:56
BLANK TEST	0.00	21:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:57
BLANK TEST	0.00	21:58

