

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

BFO

Date: 10/23/2024

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:48
INTERNAL STANDARD	VERIFIED	21:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:48
BLANK TEST	0.00	21:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:50
BLANK TEST	0.00	21:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:51
BLANK TEST	0.00	21:52

