

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

*BFO*

Date: 10/31/2024

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:13
INTERNAL STANDARD	VERIFIED	01:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:15
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:16
BLANK TEST	0.00	01:17

