

SUPERVISOR TEST

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 11/14/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:58
INTERNAL STANDARD	VERIFIED	22:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:58
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:00
BLANK TEST	0.00	23:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:01
BLANK TEST	0.00	23:02

