

SUPERVISOR TEST

BFD

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 12/11/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:30
INTERNAL STANDARD	VERIFIED	23:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:30
BLANK TEST	0.00	23:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:31
BLANK TEST	0.00	23:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:32
BLANK TEST	0.00	23:33

