

SUPERVISOR TEST

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/25/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:04
INTERNAL STANDARD	VERIFIED	22:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:06
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:07
BLANK TEST	0.00	22:08

