

SUPERVISOR TEST

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/11/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35663
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:55
INTERNAL STANDARD	VERIFIED	18:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:55
BLANK TEST	0.00	18:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:56
BLANK TEST	0.00	18:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:58
BLANK TEST	0.00	18:58

