

SUPERVISOR TEST

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 05/30/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:30
INTERNAL STANDARD	VERIFIED	21:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:31
BLANK TEST	0.00	21:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:32
BLANK TEST	0.00	21:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:33
BLANK TEST	0.00	21:34

